

Commissioning and QA of SGRT on Ring Gantry and Sim-RT installation

Mr. Sachin Kisan Rasal
Medical Physicist/RSO
Sir HN Reliance Foundation Hospital
Mumbai




ETHOS : AI Driven Adaptive Radiotherapy

Delivery :

- 6X FFF Photon, Dose rate : 800MU/min
- Delivery at 2RPM
- Novel MLC :
 - Double Bank
 - Stacked/Staggered
- Beam Stopper

Imaging:

- kVCBCT
- Imaging at 4rpm: ~ 17seconds acquisition for kV CBCT
- Large FOV up to 49cm \varnothing X 24.5cm Length
- Extended CBCT range (38.5 cm)
- Iterative CBCT available for all protocols



ETHOS : AI Driven Adaptive Radiotherapy


Not available :

- Optical Light Field
- ODI
- Hand Pendent
- Couch rotation
- Jaws
- Electron beams
- No MV imaging



C Type Linac Vs Ring Gantry Linac : SGRT

C Type Linac : Only Ceiling Mounted Camera




Ceiling Mounted Camera : Patient Setup



ALIGNRT setup

Sir H N

Each pod contains a projector, two image sensors (cameras) and a white Calibration LED



Projector - Projects a pseudo-random speckle pattern onto the patient surface which provides texture variations required for the reconstruction process

Image sensors – Acquire the raw textured data used for 3D surface reconstruction

White Calibration LED – Used for illumination during system calibration

- **Setup**
 - Three ceiling mounted camera pods
 - One central and two lateral pods
- **Surface Reconstruction:**
 - Each pod contains two camera sensors and a projector enabling real time 3D surface reconstruction
- **Registration**
 - The live surface is registered to a reference surface generating 6DOF shift information (real time deltas)
 - Frame rates up to 25 fps with AlignRT v6.0



ALIGNRT – HOW IT WORKS


3 pods project a speckled red light pattern onto patient's surface



Stereo camera pods image pattern in 3D and software reconstructs full surface



Surface matched in real time to reference image (from CT or AlignRT)




VRT _{cm}	0.02	
LNG _{cm}	0.07	
LAT _{cm}	0.02	
MAG _{cm}	0.08	
Rtn °	0.1	
Roll °	0.2	
Pitch °	-0.3	



In Bore Camera

In Bore Camera : Patient Monitoring



In Bore Camera Specification

- Three arch/Parts : Simple Locking clip system
- Dimension : 1m Diameter,
Width : 16cm,
Thickness : 3.0cm,
Weight: 2.7kg
- Material : Carbon Fiber
- Rubber Fiction mounts around the ring
- Installed at Midway : Virtual Iso and Treatment ISO



SGRT in C Type Linac vs In Bore

No Visible Speckle pattern on surface



Speckle pattern on surface




Phantom Used: Commissioning

MV Isocenter Cube Phantom




Daily Calibration Plate



Leg Phantom



Mannequin



Test Performed

- Plate Calibration
- Stability Of Camera In Bore and Ceiling Mounted
- Relative Shift accuracy
- Delta Stability and Pod Occlusion test
- Mechanical stability of In Bore Camera



Test Performed

- CBCT QA with and without Ring
- Dose measurement with and without Ring
- Integration test with ETHOS TPS
- Ambient Lighting : Alignment and Treatment



Stability of Camera

A phantom is monitored for 20 minutes

Test Name	In Bore Camera			Set up Camera		
	Maximum/Threshold	Error Read	PASS/FAIL	Maximum/Threshold	Error Read	PASS/FAIL
Thermal Stability	Drift LNG < 1.0 mm	0.20 mm	PASS	Drift LNG < 1.0 mm	0.37 mm	PASS
	Drift LAT < 1.0 mm	0.13 mm	PASS	Drift LAT < 1.0 mm	0.52 mm	PASS
	Drift VRT < 1.0 mm	0.11 mm	PASS	Drift VRT < 1.0 mm	0.29 mm	PASS
	Drift ROLL < 1.0 °	0.06 °	PASS	Drift ROLL < 1.0 °	0.08 °	PASS
	Drift PITCH < 1.0 °	0.08 °	PASS	Drift PITCH < 1.0 °	0.09 °	PASS
	Drift YAW < 1.0 °	0.04 °	PASS	Drift YAW < 1.0 °	0.09 °	PASS

Relative Shift accuracy : Tolerance (0.5mm)

In Bore Camera


Test Name	Maximum/Threshold	Error Read	PASS/FAIL
Relative Shift Accuracy Test VRT + 10 mm	VRT error < 1.0 mm	0.29 ± 0.01 mm	PASS
Relative Shift Accuracy Test LNG + 10 mm	LNG error < 1.0 mm	0.06 ± 0.04 mm	PASS
Relative Shift Accuracy Test LAT + 10 mm	LAT error < 1.0 mm	0.08 ± 0.02 mm	PASS

Relative Shift accuracy : Tolerance (0.5mm)

Set up Camera			
Test Name	Maximum/Threshold	Error Read	PASS/FAIL
Relative Shift Accuracy Test VRT + 10 mm	VRT error < 1.0 mm	0.05 ± 0.03 mm	PASS
Relative Shift Accuracy Test LNG + 10 mm	LNG error < 1.0 mm	0.13 ± 0.03 mm	PASS
Relative Shift Accuracy Test LAT + 10 mm	LAT error < 1.0 mm	0.06 ± 0.05 mm	PASS

Verify Co- Calibration of SE up and In Bore Ring Camera : Tolerance (0.5mm)

SETUP Camera	Vert : 0.1mm
Real Time Monitoring 30Sec	Lng :0.1mm
	Lat : 0.1mm
In-Bore Camera	Vert: 0.1mm
Real Time Monitoring 30Sec	Lng: 0.3mm
	Lat: 0.3mm



Delta Stability and Pod Occlusion test

Camera	Test Name	Maximum/Threshold	Error Read	PASS/FAIL
In Bore Camera	Delta Stability and Pod Occlusion Test Gantry at 0	LNG < 1.0 mm	0.14 ± 0.09 mm	PASS
		LAT < 1.0 mm	0.01 ± 0.03 mm	PASS
		VRT < 1.0 mm	0.10 ± 0.02 mm	PASS
		ROLL < 1.0 °	0.11 ± 0.01 °	PASS
		PITCH < 1.0 °	0.00 ± 0.03 °	PASS
		YAW < 1.0 °	0.05 ± 0.02 °	PASS
		MAG Std Dev < 0.2 mm	0.07 mm	PASS
Camera	Test Name	Maximum/Threshold	Error Read	PASS/FAIL
Set up Camera	Delta Stability and Pod Occlusion Test Gantry at 0	LNG < 1.0 mm	0.12 ± 0.03 mm	PASS
		LAT < 1.0 mm	0.47 ± 0.04 mm	PASS
		VRT < 1.0 mm	0.01 ± 0.03 mm	PASS
		ROLL < 1.0 °	0.02 ± 0.01 °	PASS
		PITCH < 1.0 °	0.01 ± 0.02 °	PASS
		YAW < 1.0 °	0.13 ± 0.01 °	PASS
		MAG Std Dev < 0.2 mm	0.04 mm	PASS



Ambient Lighting : Alignment and Treatment

Camera	Test Name	Maximum/Threshold	Error Read	PASS/FAIL
In Bore Camera	Ambient Lighting during Patient Alignment and Treatment	Drift LNG < 0.2 mm	0.02 mm	PASS
		Drift LAT < 0.2 mm	0.01 mm	PASS
		Drift VRT < 0.2 mm	0.01 mm	PASS
		Drift ROLL < 0.2 °	0.00 °	PASS
		Drift PITCH < 0.2 °	0.01 °	PASS
		Drift YAW < 0.2 °	0.00 °	PASS
Camera	Test Name	Maximum/Threshold	Error Read	PASS/FAIL
Set up Camera	Ambient Lighting during Patient Alignment and Treatment	Drift LNG < 0.2 mm	0.01 mm	PASS
		Drift LAT < 0.2 mm	0.00 mm	PASS
		Drift VRT < 0.2 mm	0.01 mm	PASS
		Drift ROLL < 0.2 °	0.00 °	PASS
		Drift PITCH < 0.2 °	0.00 °	PASS
		Drift YAW < 0.2 °	0.00 °	PASS

Accuracy and Reproducibility

- **Accuracy**

$\Delta Vrt(cm)$	$\Delta Long(cm)$	$\Delta Lat(cm)$
0.01	0.02	0.01

- **Reproducibility was 0.02mm**



Our Initial QA Result of Adaptive with SGRT Ring

No Fraction	Gamma Pass%
Pre Treatment	
1	98.6
2	97.8
3	98.7
4	97.5
5	98.1
6	98
7	98.8
8	95.4
9	96.4
10	95.2
11	96.5
12	95.5
13	96
14	96.4
15	96.2
16	96.1
17	95.5
18	95.8
19	95
20	95.9
21	95.5
Avg	96.61

No Fraction	Gamma Pass%
Pre Treatment	
1	96.3
2	98.3
3	99.4
4	97.9
5	98.7
6	98.5
7	99.4
8	99
9	98.4
10	98.5
11	98.8
Avg	98.47

No Fraction	Gamma Pass%
Pre Treatment	
1	98.9
2	96.5
3	98.3
4	98.4
5	98.9
6	98.8
7	98.5
8	98.7
9	97.5
10	98.5
11	97.8
Avg	98.25

No Fraction	Gamma Pass%
Pre Treatment	
1	97.2
2	97.2
3	97.3
4	97.4
5	95
6	95.4
7	95.2
8	95.2
Avg	96.24

No Fraction	Gamma Pass%
Pre Treatment	
1	96.6
2	95.3
3	96.6
4	98.8
5	97.8
6	96.9
Avg	97.00



Initial Experience-CBCT and SGRT Shift

Breast		176 #
$\Delta\text{vrt}(\text{cm})$	$\Delta\text{long}(\text{cm})$	$\Delta\text{lat}(\text{cm})$
0.23	0.47	0.17
0.38	0.20	0.21
0.31	0.34	0.19
0.46	0.30	0.32
0.23	0.37	0.40
0.18	0.32	0.27
0.30	0.27	0.33
0.20	0.22	0.17
0.32	0.58	-0.02
0.12	0.00	0.11
-0.22	-0.48	0.27

Prostate		250 #
$\Delta\text{vrt}(\text{cm})$	$\Delta\text{long}(\text{cm})$	$\Delta\text{lat}(\text{cm})$
0.08	0.07	0.06
0.09	0.03	0.13
0.09	0.05	0.03
-0.16	-0.32	0.04
-0.24	-0.34	-0.45
0.52	-0.34	0.02
0.27	0.02	-0.04
0.12	0.06	0.15
-0.02	0.08	0.38
-0.01	0.20	0.28
0.12	0.08	0.11



Frequency of QA

TABLE 5 Summary of routine QA tests to be performed daily, monthly, annually as specified in Table II of AAPM's Task Group 147¹

Frequency	Test category	Methods	Accuracy
Daily	Safety	Check interlocks and clear FOV for all mounted cameras	Pass
	Static localization	Daily QA phantom positioned at isocenter and can track movement to isocenter from offset	2 mm
Monthly (in addition to daily tests):	Safety	Machine interface: gating termination, couch motion communication	Functional
	Static localization	Localization test based on radiographic analysis (i.e., hidden target)	2 mm 1 mm for SRS/SBRT
	Dynamic localization	Motion table or manual couch motion of monthly phantom by known distances	2 mm or less as per manufacturer spec.
Annually (in addition to all monthly tests)	Safety	Test/reset buttons, backup power supply, and emergency-off switches	Pass
		System mounting brackets (all cameras are secure)	Pass
	Integrity	Check camera settings if accessible	Unchanged from previous
	Stability (drift/reproducibility)	Drift measurement (over at least 1 h)	<2 mm over 1 h
		Reproducibility of localization	<1 mm after stabilizing
	Static localization (extensive)	Complete end-to-end test (including data transfer check of localization accuracy, etc.)	<2 mm from isocenter <1 mm for SRS/SBRT
		Translation and rotation auto correct over a clinical range of motion	<2 mm from isocenter
	Dynamic (gating system)	Using a motion phantom/check of gating system radiation dosimetry accuracy.	< 2% (per TG-142)
	Data transfer	From all systems in use	Functional



Limitation

- MMI : Motion Management Interface not available, No auto Beam Hold
- No Direct Integration to ETHOS
- Re-calibration of Ring Post Servicing of ETHOS



Conclusion

- This Acceptance test of In Bore camera proved that overall performance is comparable to ceiling mounted camera.
- In Bore camera SGRT system sufficiently verified for clinical use.



Sim RT :Installation and Commissioning



Sim-RT

- What is Sim-RT?
- Where Sim-RT?
- Why Sim-RT?



Guideline for Surface Guided Radiotherapy

ESTRO -ACROP: Guideline on Surface-guided radiotherapy AAPM Task Group-302:Surface-guided radiotherapy

Radiotherapy and Oncology 173 (2022) 188–196



Contents lists available at ScienceDirect

Radiotherapy and Oncology

journal homepage: www.thegreenjournal.com



Guidelines

ESTRO-ACROP guideline on surface guided radiation therapy

P. Freisleder^{a,1,*}, V. Batista^{b,c}, M. Öllers^d, M. Buschmann^e, E. Steiner^f, M. Kügele^g,
F. Fracchiolla^h, S. Corradini^a, M. de Smetⁱ, F. Moura^j, S. Perryck^k, F. Dionisi^l, D. Nguyen^m,
C. Bertⁿ, J. Lehmann^{o,p,q}



^aDepartment of Radiation Oncology, LMU University Hospital, Munich; ^bDepartment of Radiation Oncology, Heidelberg University Hospital; ^cHeidelberg Institute of Radiation Oncology (HIRO), Heidelberg; ^dDepartment of Radiotherapy, MAASTRO, Maastricht, The Netherlands; ^eDepartment of Radiation Oncology, Medical University of Vienna/AKH Wien; ^fInstitute for Radiation Oncology and Radiotherapy, Landeskrankenhaus Wiener Neustadt; ^gDepartment of Hematology, Oncology and Radiation Physics, Skåne University Hospital, Lund, Sweden; ^hAzienda Provinciale per i Servizi Sanitari (APSS) Protontherapy Department, Trento, Italy; ⁱDepartment of Medical Physics & Instrumentation, Institute Verbeeten, Tilburg, The Netherlands; ^jHospital CUF Descobertas, Department of Radiation Oncology, Lisbon, Portugal; ^kDepartment of Radiation Oncology, University Hospital Zürich, Switzerland; ^lDepartment of Radiation Oncology, IRCCS Regina Elena National Cancer Institute, Rome, Italy; ^mCentre de Radiothérapie de Mâcon, France; ⁿDepartment of Radiation Oncology, Universitätsklinikum Erlangen, Friedrich-Alexander-Universität Erlangen-Nürnberg, Germany; ^oRadiation Oncology Department, Calvary Mater Newcastle; ^pSchool of Information and Physical Sciences, University of Newcastle; and ^qInstitute of Medical Physics, University of Sydney, Australia

ARTICLE INFO

Article history:
Received 24 May 2022
Accepted 26 May 2022
Available online 31 May 2022

ABSTRACT

Surface guidance systems enable patient positioning and motion monitoring without using ionising radiation. Surface Guided Radiation Therapy (SGRT) has therefore been widely adopted in radiation therapy in recent years, but guidelines on workflows and specific quality assurance (QA) are lacking. This ESTRO-ACROP guideline aims to give recommendations concerning SGRT roles and responsibilities and high-

Received: 3 October 2021 | Revised: 26 December 2021 | Accepted: 5 February 2022

DOI: 10.1002/ro.1932

AAPM SCIENTIFIC REPORT

MEDICAL PHYSICS

AAPM task group report 302: Surface-guided radiotherapy

Hania A. Al-Hallaq¹ | Laura Cerviño² | Alonso N. Gutierrez³ |
Amanda Havnen-Smith⁴ | Susan A. Higgins⁵ | Malin Kügele^{6,7} | Laura Padilla⁸ |
Todd Pawlicki⁸ | Nicholas Remmes⁹ | Koren Smith¹⁰ | Xiaoli Tang¹¹ |
Wolfgang A. Tome¹²




Installation and Commissioning

- *“Installation and commissioning is part of the QA Program and it includes testing a system capabilities and verifying its accuracy/precision in all clinically relevant scenario.”*
- Steps:
 1. Installation Components.
 2. Acceptance/QA Test Performed.



Main Components Used in Installation of Sim-RT

Camera Mount



All these items are pre-assembled from manufacturing
--Outer and Inner mount set to length at manufacturing




Main Components Used in Installation of Sim-RT

Celling Plate and Horizon Camera




Main Components Used in Installation of Sim-RT

- Real time coaching device and Calibration Plate




Camera Layout

CT Simulation Room Plan View



CT Simulation Room Sectional View






Commissioning and QA Test

- The following tests are conducted and validated the Sim-RT system:
 - System Calibration(Monthly QA)
 - Daily QA
 - Validation Test




Monthly Calibration

- Step 1. Place the calibration plate on the couch at bore centre.
- Step 2. Open SimRT using your account.
- Step 3. Click on the System **Calibration** Icon.
- Step 4. Follow the instructions on the screen and click “Start Calibration” button.
- Step 5. Click the landmark points, 1-4 as shown in image




Monthly Calibration

- Step 6. Next click “Continue”
- Step 7. The software will ask to advance the couch by 25cm into the bore.
Advance the CT couch and then click “Continue”.
- Step 8. Select the landmark number corner points in order 1-4 as shown below.
- Step 9. Click “Apply” and then click “OK”. Monthly Calibration is completed.




DAILY QA

- The Daily QA verifies there has been no camera pod movement since the last monthly calibration.
- *Note: Daily QA must be performed after any Monthly Calibration for Sim-RT to function properly.*
- Step 1. Place the calibration plate on the couch at bore centre.(Same As Monthly QA)
- Step 2. Open SimRT using your account.
- Step 3. Click on the System **Daily QA Icon**.
- Step 4. Click on “Start Daily QA”



DAILY QA

- Step 5. The software will ask to advance the couch by 25cm into the bore. Advance the CT
- couch and then click "Continue". And Observe the shift



SIM RT Validation test



- Following test must be run in order to shown below.
- 1. Shift Accuracy:** The Average displacement of each three shift computed from VXP files and compare to initial one and difference must be **< 0.5mm** .
 - 2. Monitoring Stability:** Standard deviation from VXP file **<0.2mm**.
 - 3. Frame Rate:** frame rate from VXP file **> 25fps**.



SIM RT Validation: Create a VXP file(5min)

Setup:

- Set a SimRT Spacer at its lowest height(7cm) and aligned with room laser,
- Place a ACO plate on the Spacer and aligned with room laser(see following picture.)



SIM RT Validation: Create a VXP file

- Create A test Patient(Patient Name Sim RT Validation)
- Choose a “DIBH(3D Helical) Protocol”
- Capture a Surface of ACO plate.
- Position patch on centre of plate and monitor for 20sec.
- Stop monitoring and revive and export the file .
- Repeat for 8cm and 9cm ,export file .
- Do analysis of file in “Accuracy Validation tool ”



SIM RT Validation: Create a VXP file:Result

Accuracy Validation Test (SimRT only)

System validation analysis test (SimRT only)

☒ Pass ☐ Fail ☐ N/A

Shift Accuracy: The average displacement for each of the three shifts computed from VXP files and compared to the ground truth.

Pass

0mm shift error <0.5mm	<input type="text" value="0.06"/>	-	+
10mm shift error <0.5mm	<input type="text" value="0.20"/>	-	+
20mm shift error <0.5mm	<input type="text" value="0.07"/>	-	+

Monitoring Stability: Standard deviation of the displacement values computed from VXP file.

Pass

Standard deviation <0.2mm	<input type="text" value="0.04"/>	-	+
---------------------------	-----------------------------------	---	---

Frame rate: Frame Rate computed from VXP file

Pass

Frame rate > 25fps	<input type="text" value="35"/>
--------------------	---------------------------------



Advantages

- Tracking Point Selection inside the CT Room
- The Ceiling Mounted Camera allows for no physical marker, block or belt needed
 - no physical distraction for patients
- no skin deformation, no inaccurate block placement, and nothing to clean.
- completely non-invasive and non-ionising motion monitoring.
- Real Time Coach Display for coaching and monitoring breathing pattern
- Advance Software for tracking and monitoring
- quick and efficient setup
- Doesn't require any additional equipment to be setup for each patient



Thank You

